



U.S. DEPARTMENT OF HEALTH AND HUMAN SERVICES  
**FOOD AND DRUG ADMINISTRATION**



*Certificate of Class I*

*Awarded to*

**Megagen Implant Co., Ltd.**

377-2 Gyocheon-Ri, Jain-Myeon, Gyeongsan-Si, South Korea

**Listing Number: D061503**

**Registration Number: 3005554774**

This is to certify that Restome Kit and Megatome Kit have been registered with the U.S. FDA and it is safe for marketing in the United States of America.

**Indications for Use:** Restome Kit and Megatome Kit is a dental hand device intended to perform bone expansion, bone compaction, bone elevation and sinus graft in general dentistry and oral surgery procedures.

**Product Class:** Chisel, Osteotome, Surgical

**Product Code:** EMM

**Subsequent Code:** HXL

**Trade Name:** **Restome Kit, Megatome Kit**

**Regulation Number:** 21 CFR 872.4565

**Regulatory Class:** Class I

**FDA Registered Date:** November 14, 2008

Presented by: Official Correspondent, Consultant and U.S. FDA Designated Agent



Kodent Inc.

13340 E. Firestone Blvd. Suite J, Santa Fe Springs, CA 90670 USA

*Jung Bae Bang*