



Certificate of Class I Awarded to

Megagen Implant Co., Ltd.

377-2 Gyochun-ri, Jain-myun, Gyeongsan, Gyeongbuk, KOREA, REPUBLIC OF

Listing Number: D097207 Registration Number: 3005554774

This is to certify that Impression Tray has been registered with the U.S. FDA and it is safe for marketing in the United States of America.

Indications for Use: IMPRESSION TRAY is a dental impression tray for taking impression

on edentulous jaws or teeth of implant patient. It is used for pick-up

impression (open-tray) or transfer impression (close tray).

Product Class: Tray, Impression, Performed

Product Code: EHY

Trade Name: Impression Tray

Regulation Number: 21 CFR 872.6880

Regulatory Class: Class I

FDA Registered Date: April 15, 2010

Presented by: Official Correspondent, Consultant and U.S. FDA Designated Agent



Kodent Inc.

13340 E. Firestone Blvd. Suite J, Santa Fe Springs, CA 90670 USA

Dr. Steve Chang