



U.S. DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FOOD AND DRUG ADMINISTRATION



*Certificate of Class I*

*Awarded to*

**Megagen Implant Co., Ltd.**

377-2 Gyocheon-Ri, Jain-Myeon, Gyeongsan-Si, South Korea


**Listing Number: E220672**

**Registration Number: 3005554774**

This is to certify that Internal Sinus Kit has been registered with the U.S. FDA and it is safe for marketing in the United States of America

<b>Indications for Use:</b>	Internal Sinus Kit is a rotary device intended to be attached to a dental hand piece to drill holes in teeth to secure cast or performed pins to retain operative dental appliances.
<b>Product Class:</b>	Accessories, Implant, Dental, Endosseous
<b>Product Code:</b>	NDP
<b>Trade Name:</b>	Internal Sinus Kit
<b>Regulation Number:</b>	21 CFR 872.3980
<b>Regulatory Class:</b>	Class I
<b>FDA Registered Date:</b>	Nov 24, 2008

Presented by: Official Correspondent, Consultant and U.S. FDA Designated Agent



Kodent Inc.

13340 E. Firestone Blvd. Suite J, Santa Fe Springs, CA 90670 USA

*Jung Bae Bang*